

湖北省计量测试技术研究院

Huber Institute of Measurement and Testing Technology

地址 (Add) : 湖北省武汉市东湖新技术开发区
茅店山中路二号 邮编 (Post Code) : 430223
网址 (Web site) : <http://www.hbjl.gov.cn> 电话 (Tel) : 027-81925136
传真 (Fax) : 027-81925137

中国校准服务联合体
China United Calibration Service

CUC

REGISTRATION NO. 002

校准证书 CALIBRATION CERTIFICATE

证书编号:
Certificate No.

委托方名称 中国能源建设集团湖南火电建设有限公司

Customer

委托方地址 湖南省长沙市天心区雀园358号

Address

器具名称 干式试验变压器

Name Of Sample

制造厂商 武汉特高压电力科技有限公司

Manufacturer

型号规格 GTB-5kVA/50kV

Model/Type

器具编号 1608078

No Of Sample

证书专用章

Stamp



校准日期

Calibration date

建议再校日期

Recommended recal.date

2024年 10月 06日

Y M D

2025年 10月 05日

Y M D

证书批准人

Approved by

核验员

Checked by

校准员

Calibrated by

耿宵
田天

蔡建军

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基标准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.

- 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

- 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

- 本次校准的技术依据（名称、代号）

Reference documents for the Calibration (Name、Code)

参照: DL/T848.2-2004 高压试验装置通用技术条件 第2部分: 工频高压试验装置

- 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称
Name of Equipment
数字高压表

型号/编号
Model/Serial No.
FRC-150M/11090603

证书号/有效期
Certificate No./Due Date
2024DW2380054/2025-10-05

计量测试技术
证书骑缝章

- 校准环境条件

Environmental condition on the Calibration

温度: 25.6 °C

相对湿度: 65 %

其它:

Temperature

R.H.

Others

气压:

地点: 现场

原始记录编号:

Pressure

Place

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.

未经本院许可, 不得部分复制本证书。

校准数据/结果

Data/Results of Calibration

高电压测量: AC				
被检显示值 (kV)	实测标准值 (kV) AC	测量不确定度: $U(kV), k=2$	实测标准值 (kV) DC	测量不确定度: $U(kV), k=2$
10	10.70	0.10	11.20	0.10
20	20.25	0.10	20.63	0.10
30	29.67	0.15	29.94	0.15
40	39.20	0.20	39.08	0.20
50	49.12	0.25	48.09	0.25
60	----	----	57.60	0.30
70	----	----	67.18	0.35

以下空白

湖北省

研究院